

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

System: Skyrad NIP

Site: Darwin

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 450)X10-6	Offset	Date of change UTC	Time Start UTC	Time End UTC
NIP	33551E6	Skyrad	24 May 2004	8.04		3 Sept 2004	03:00	04:00

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 450)X10-6	Offset	Copy Config. Logger to PC
NIP	31875E6	Skyrad	3 Sept 2004	8.4442		yes

Replacement completed by: Troy Culgan & Bill Porch

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP

Other old part disposition (explain): _____

Other NOTES: 33551E6 was installed during SV0307, however, when the new Campbell loggers were installed in May 2004, the Cal Gain Coefficient was erroneously entered as 8.62. This was found and correct on 3 Sept. before the instrument was replaced.