

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0508D

System: Skyrad PIRD2

Site: Darwin

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
PIR	31391	Skyrad	9/2/2004	-7.32 K0 .26076 K1 1.008 K2 (-2.782) K3	na	8/5/2005	00:00	1:30

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
PIR	33059F3	Skyrad	8/5/2005	-7.62 K0 .25334 K1 1.013 K2 (-2.427) K3	na	X

Replacement completed by: Gary Eckert, Mike Alsop, Bill
Porch

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other comments: