

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0611N_cal

System: SMET

Site: Nauru

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
PTB201A	P0830005	Barometer	V060901.00	na	na	1/Dec/06	01:46	04:30

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
05106/08234	505303	Barometer	V060901.00	na	na	

Replacement completed by: Troy Culgan

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: Darwin for assessment

Other old part disposition (explain):

Other comments: Barometer changed out as it was outside calibration parameters