

# ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

System: Smet – Wind High \_\_\_\_\_

Site: ARCS2 \_\_\_\_\_

## OLD Configuration

Part Number	Serial Number	Location	Configuration Version	Slope	Intercept	Date UTC	Time Start UTC	Time End UTC
Anemometer nonellor	49705 76843	SMet	N/A	0.09663	0.35	3 Mar 05	0300	0430

## NEW Configuration

Part Number	Serial Number	Location	Configuration Version	Slope	Intercept	Copy Config. Logger to PC	Save to EPROM(E)	Update Version #
Anemometer nonellor	31103 58318	SMet	N/A	0.09805	0.11	N/A	N/A	N/A

Replacement completed by: Nauru Observers \_\_\_\_\_

## **NOTES:**

Old Part remains on site as spare: No

Old part packed for shipment to: SGP \_\_\_\_\_

Other old part disposition (explain): \_\_\_\_\_

Other NOTES: Returned for repair/calibration