

ARCS PROCEDURE: Author: P. Johnston	ISS EMERGENCY SHUTDOWN PROCEDURE	PRO(ISS)-003.000 April 13, 1999 Page 1 of 1
--	-------------------------------------	---

ISS Emergency Shutdown Procedure

I. Purpose:

This document describes the emergency shutdown procedure for the Nauru NOAA ISS. This is a small Doppler radar that measures the winds above the station. It also has a small automatic surface meteorological station.

II. Cautions and Hazards:

- None.

III. Requirements:

- None.

IV. Procedure:

A. Steps:

1. If there is a reason to turn the equipment off quickly, turn off the power to the system. This is done easiest by turning off the top breaker inside electric panel just inside the door.
2. The power can also be turned off at the back of the container next to the meter.

V. References:

None.

VI. Attachments:

None.