

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0511M

System: Skyrad PIRD2

Site: Manus

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
PIR	31302F3	Skyrad	11/04/2004	-5.41 K0 .27536 K1 1.007 K2 (-2.405) K3	na	11/15/2005	03:00	04:15

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
PIR	30167F3	Skyrad	11/15/2005	-4.82 K0 .21545 K1 1.005 K2 (-2.139) K3	na	X

Replacement completed by: Troy Culgan

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other comments: