

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV0611N_cal

System: Gndrad IRT

Site: Nauru

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
KT19.85	863	GRDRAD	V061202	n/a	na	4/Feb/2007	23:40	23:42

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
KT19.85	865	GRDRAD	V070205.0	n/a	na	x

Replacement completed by: Franklin Teimitsi

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP for repair

Other old part disposition (explain):

Other comments: T.low -50C, T.end +200C, Tr 3.0, Volts 0..10